

# CLEANBRUSH® ENDO-XL

56

CLEANING BRUSH FOR RIGID INSTRUMENTS



The **CLEANBRUSH® ENDO-XL** cleaning brushes are intended for cleaning the channels of rigid instruments before sterilisation, such as surgical trocars.

They are also suitable for cleaning rigid endoscopes used in urology, gynaecology and ENT.

## AVAILABLE REFERENCE(S)

Reference	Brush diameter (mm)	Number of brushes	Sheath diameter (mm)	Recommended diameter for channel to be cleaned	Length (cm)	Packaging
AN3435506	5.5	1	3.3	From 4.1 to 5 mm	60	20
AN3438006	8	1	3.3	From 5.1 mm to 7 mm	60	20
AN3431106	11	1	3.3	From 7.1 to 9 mm	60	20
AN3431406	14	1	3.3	From 9.1 mm to 12 mm	60	10
AN3431806	18	1	3.3	From 12.1 to 16 mm	60	10
AN2575006	5 11	2	3.3	From 2.7 to 4.5 mm From 6 to 10 mm	60	10



## ABOUT THIS MEDICAL DEVICE

1 <sup>st</sup> release date	1999
Single use	Yes
Class	I
Sterile	No (ethylene oxide decontamination)
Contains latex	No
Lifetime	5 years
Certification	CE
Packaging	Box of 10 or 20 units in non-sterile wrapping according to the reference
Storage	In its original box in a cool and dry place

PRINCE MEDICAL SAS  
 ZA La Sente Du Moulin  
 64 rue Gutenberg - 60530 Ercuis - France  
[www.prince-medical.com](http://www.prince-medical.com)  
 Tel: +33 (0)3 44 26 54 76  
[prince-medical@omerin.com](mailto:prince-medical@omerin.com)

**PRINCE MEDICAL**  
 DEVICES

The information provided on this data sheet is indicative and subject to change without notice. PRINCE MEDICAL SAS shall not be held responsible for any incidents resulting from inappropriate use, failure to comply with the rules and regulations in force, or failure to follow the instructions for use provided. Medical devices sold by PRINCE MEDICAL SAS should only be used by a duly qualified practitioner or authorised person.

® Registered trademark of PRINCE MEDICAL SAS. Drawings and photos are not contractually binding. Reproduction is prohibited without the prior consent of PRINCE MEDICAL SAS.